

## Representative Reactivities Table



### Optitrol NAT Triplex

Positive Control

Version: 2

Date: July 2nd, 2022

Analyte	Manufacturer	Assay Name	Unit	Mean	SD	Samples	Labs	Representative Reactivity
HCV	Roche	cobas® MPX Multiplex HIV, HCV & HBV nucleic acid test for use on cobas® 6800/8800 Systems	Ct	36.3	0.4	30	3	positive
HBV			Ct	32.6	0.3	30	3	positive
HIV			Ct	34.2	0.3	30	3	positive

The stated mean values were obtained by testing one product batch. They show possible representative values but no target values.  
Test results may vary among used instruments, assays, reagent lots, and different laboratory protocols.