

Representative Reactivities Table



Optitrol NAT CMV

Positive Control

Version: 1

Date: July 05, 2021

Analyte	Manufacturer	Assay Name	Extraction method	Unit	Mean	SD	Samples	Labs	Representative Reactivity
CMV	Altona	Altona Real Star CMV PCR Kit 1.0	QIAamp Mini Elute Virus Spin Kit	IU/mL	1544	43.3	10	1	positive
	Altona	Altona Real Star CMV PCR Kit 1.0	Roche Magna Pure 96 DNA/Viral NA SV 2.0 Kit	IU/mL	2280	29.0	10	1	positive
	Abbott	Abbott Real Time CMV Amp Kit	Abbott Preparation System DNA	IU/mL	1981	37.1	10	1	positive

The stated mean values were obtained by testing one to three product batches. They show possible representative values but no target values. Test results may vary among used instruments, assays, reagent lots, and different laboratory protocols.