

Representative Reactivities Table



Optitrol NAT HIV-2

Positive Control

Version: 1

Date: April 5, 2022

Analyte	Manufacturer	Assay Name	Unit	Mean	SD	Samples	Labs	Representative Reactivity
HIV-2	Roche	cobas® HIV-1/HIV-2 Qualitative for use on cobas® 5800/6800/8800 Systems	IU/mL	136.1	25.4	10	1	positive
	Grifols	Procleix® Ultrio Elite Assay Procleix Panther System	IU/mL	120.5	10.5	10	1	positive

The stated mean values were obtained by testing one product batch. They show possible representative values but no target values. Test results may vary among used instruments, assays, reagent lots, and different laboratory protocols.