

Representative Reactivities Table
QC Release Values
Optitrol Multiallergy Food



DM22220 Exp. 2024-12

(Included in Lot DM22222, DM22223 and DM22224)

V001 Date: 09.01.2023

positive
grayzone / negative

Code	Name	MEDIWISS, AlleisaScreen®		
		Mean value [IU/L]	Acceptable range [IU/L]	Acceptable range [CAP]
f1	Egg white	0.0	0.0 - 0.69	0 - 1
f2	Milk	0.0	0.0 - 0.69	0 - 1
f3	Cod fish	0.0	0.0 - 0.69	0 - 1
f41	Salmon	0.0	0.0 - 0.69	0 - 1
f4	Wheat flour	31.3	3.5 - 100	3 - 5
F5	Rye flour	24.7	3.5 - 100	3 - 5
f12	Pea/Bean	0.3	0.0 - 3.49	0 - 2
f13	Peanut	5.4	0.7 - 49.9	2 - 4
f14	Soybean	2.8	0.35 - 17.49	1 - 3
f17	Hazelnut	9.2	0.7 - 49.9	2 - 4
f23	Crab	0.0	0.0 - 0.69	0 - 1
f24	Shrimp	0.0	0.0 - 0.69	0 - 1
f25	Tomato	22.7	3.5 - 100	3 - 5
f31	Carrot	17.0	3.5 - 100	3 - 5
f33	Orange	10.3	0.7 - 49.9	2 - 4
f35	Potatoe	22.0	0.7 - 49.9	2 - 4
f48	Onion	8.4	0.7 - 49.9	2 - 4
f49	Apple	5.8	0.7 - 49.9	2 - 4
f59	Cuttlefish	0.0	0.0 - 0.69	0 - 1
f207	Mussel	0.0	0.0 - 0.69	0 - 1
f75	Egg yolk	0.0	0.0 - 0.69	0 - 1
f76	alpha-Lactglobulin	0.0	0.0 - 0.69	0 - 1
f77	beta-Lactglobulin	0.0	0.0 - 0.69	0 - 1
f78	Casein	0.0	0.0 - 0.69	0 - 1
f84	Kiwi	11.8	0.7 - 49.9	2 - 4
f85	Celery	29.8	3.5 - 100	3 - 5
f91	Mango	13.7	0.7 - 49.9	2 - 4
f92	Banana	2.7	0.35 - 17.49	1 - 3
f218	Pepper	13.5	0.7 - 49.9	2 - 4
f256	Walnut	15.2	0.7 - 49.9	2 - 4
CCD1	Bromelain	> 100	17.5 - > 100	4 - 6
CCD2	Horseradish peroxidase	49.3	17.5 - > 100	4 - 6
CCD3	Ascorbatoxidase	> 100	17.5 - > 100	4 - 6

The stated mean values were obtained by testing the specifically indicated QC production lot. These values are indicative only, and therefore test results may vary among different laboratory protocols and especially by using different assays or assay reagent lots.