

Representative Reactivities Table



Optitrol HIVp24

Positive Control

Version: 2

Date: September 20, 2021

Analyte	Manufacturer	Assay Name	Method	Unit	Mean	SD	Samples	Labs	Representative Reactivity
HIVp24	Abbott	Architect HIV Ab/Ag Combo	CMIA	S/CO	4.09	0.13	10	2	positive
HIVp24	DiaSorin	LIAISON Murex HIV Ab/Ag	CLIA	S/CO	2.02	0.09	10	2	positive
HIVp24	Roche	Elecsys HIV Combi PT	ECLIA	COI	2.80	0.30	10	2	positive
HIVp24	Siemens	Centaur ADVIA HIV Ag/Ab Combo	ChLIA	COI	2.46	0.14	10	2	positive
HIVp24	Siemens	Atellica HIV Ag/Ab	ChLIA	Index	2.49	0.07	10	1	positive

The stated mean values were obtained by testing one product batch. They show possible representative values but no target values.
Test results may vary among used instruments, assays, reagent lots, and different laboratory protocols.