

Representative Reactivities Table



Optitrol Red

Positive Control

Version: 3

Date: January 26, 2022

Analyte	Manufacturer	Assay Name	Platform	Method	Unit	Mean	SD	Samples	Labs	Representative Reactivity
Anti-HCV IgG	Roche	Elecsys Anti-HCV II	cobas e analysers	ECLIA	COI	6.06	0.09	10	2	positive
Anti-HIV 1 IgG	Roche	Elecsys HIV Combi PT	cobas e analysers	ECLIA	COI	5.77	0.07	10	2	positive
HBsAg	Roche	Elecsys HBsAg II	cobas e analysers	ECLIA	COI	4.15	0.77	10	2	positive
Anti-HBc IgG	Roche	Elecsys Anti-HBc II	cobas e analysers	ECLIA	COI	0.15	0.01	10	2	positive
Anti-HTLV I IgG	Roche	Elecsys HTLV I/II	cobas e analysers	ECLIA	COI	2.45	0.02	10	1	positive
Anti- <i>Treponema pallidum</i> IgG	Roche	Elecsys Syphilis	cobas e analysers	ECLIA	COI	5.46	0.15	10	1	positive

The stated mean values were obtained by testing one product batch. They show possible representative values but no target values.

Test results may vary among used instruments, assays, reagent lots, and different laboratory protocols. **The mean values above were measured on cobas e 411 and on cobas e 602.**