Target Values / Zielwerte
Optitrol Multiallergy Food

Table 1

April 15, 2019

Standardparameter of the lot DM 19034 Exp. 2020-07

positive	grayzone	negative	/ = not tested

Code	Name	AllergyScreen® [kU/L] DiaMex Value	AllergyScreen® [kU/L] Target Value	AllergyScreen® [CAP] Target Value	ImmunoCAP® [kU/L] DiaMex Value	ImmunoCAP® [kU/L] Target Value	ImmunoCAP® [CAP] Target Value	Polychek® [kU/L] DiaMex Value	Polychek® [kU/L] Target Value	Polychek® [CAP] Target Value	TEST TAP® [kU/L] DiaMex Value	TEST TAP® [kU/L] Target Value	TEST TAP® [CAP] Target Value
f1	Egg white	< 0,35	< 0,35 - 0,69	0 - 1	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0
f2	Milk	< 0,35	< 0,35 - 0,69	0 - 1	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0
f3	Cod fish	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0
f4	Wheat flour	31,23	3,5 - 100	3 - 5	1,9	< 0,35 - 3,49	0 - 2	2,9	< 0,35 - 3,49	0 - 2	6,0	0,35 - 17,49	1 - 3
F5	Rye flour	24,87	3,5 - 100	3 - 5	1,8	< 0,35 - 3,49	0 - 2	3,3	0,35 - 17,49	1 - 3	/	/	/
f9	Rice	/	/	/	2,5	< 0,35 - 3,49	0 - 2	11	0,69 - 49,9	2 - 4	/	/	/
f10	Sesame	/	/	/	3,3	0,35 - 17,49	1 - 3	2,2	< 0,35 - 3,49	0 - 2	/	/	/
f12	Pea/Bean	1,01	< 0,35 - 3,49	0 - 2	1,9	< 0,35 - 3,49	0 - 2	/	/	/	1,6	< 0,35 - 3,49	0 - 2
f13	Peanut	4,43	0,35 - 17,49	1 - 3	3,9	0,35-17,49	1-3	5,7	0,35 - 17,49	1 - 3	6,9	0,35 - 17,49	1 - 3
f14	Sojbean	4,03	0,35 - 17,49	1 - 3	7,1	0,35 - 17,49	1 - 3	3,8	0,35 - 17,49	1 - 3	6,1	0,35 - 17,49	1 - 3
f17	Hazelnut	15,37	0,7 - 49,9	2 - 4	24,2	0,7 - 49,9	2 - 4	8,7	0,35 - 17,49	1 - 3	14	0,69 - 49,9	2 - 4
f20	Almond	/	/	/	2,2	< 0,35 - 3,49	0 - 2	2,7	< 0,35 - 3,49	0 - 2	/	/	/
f23	Crab	< 0,35	0	0	1	< 0,35 - 3,49	0 - 2	/	/	/	6,4	0,35 - 17,49	1 - 3
f24	Shrimp	< 0,35	0	0	0,78	< 0,35 - 3,49	0 - 2	< 0,35	0	0	< 0,35	0	0
f25	Tomato	19,3	0,7 - 49,9	2 - 4	2,8	< 0,35 - 3,49	0 - 2	9,3	0,35 - 17,49	1 - 3	21	0,69 - 49,9	2 - 4
f31	Carrot	23,6	3,5 - 100	3 - 5	4,6	0,35 - 17,49	1 - 3	7,9	0,35 - 17,49	1 - 3	/	/	/
f33	Orange	8,3	0,35 - 17,49	1 - 3	< 0,35	0	0	/	/	/	/	/	/
f35	Potatoe	26,13	3,5 - 100	3 - 5	3	0,35 - 17,49	1 - 3	23	0,69 - 49,9	2 - 4	/	/	/
f41	Salmon	< 0,35	0	0	/	/	/	/	/	/	/	/	/
f48	Onion	6,67	0,7 - 49,9	2 - 4	1,8	< 0,35 - 3,49	0 - 2	/	/	/	22	0,69 - 49,9	2 - 4
f49	Apple	10,97	0,7 - 49,9	2 - 4	7,1	0,35 - 17,49	1 - 3	5,8	0,35 - 17,49	1 - 3	/	/	/

Target Values / Zielwerte
Optitrol Multiallergy Food

Table 2

Standardparameter of the lot DM 19034 Exp. 2020-07

DiaMex GmbH

April 15, 2019

positive	grayzone	negative	/ = not tested

Code	Name	AllergyScreen® [kU/L] DiaMex Value	AllergyScreen® [kU/L] Target Value	AllergyScreen® [CAP] Target Value	ImmunoCAP® [kU/L] DiaMex Value	ImmunoCAP® [kU/L] Target Value	ImmunoCAP® [CAP] Target Value	Polychek® [kU/L] DiaMex Value	Polychek® [kU/L] Target Value	Polychek® [CAP] Target Value	TEST TAP® [kU/L] DiaMex Value	TEST TAP® [kU/L] Target Value	TEST TAP® [CAP] Target Value
f59	Cuttlefish	< 0,35	0	0	/	/	/	/	/	/	/	/	/
f75	Egg yolk	< 0,35	< 0,35 - 0,69	0 - 1	< 0,35	0	0	< 0,35	0	0	/	/	/
f76	alpha-Lactglobulin	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0
f77	beta-Lactglobulin	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0
f78	Casein	< 0,35	< 0,35 - 0,69	0 - 1	< 0,35	0	0	< 0,35	0	0	< 0,35	0	0
f83	Chicken meat	/	/	/	/	/	/	< 0,35	0	0	/	/	/
f84	Kiwi	9,8	0,7 - 49,9	2 - 4	4,2	0,35 - 17,49	1 - 3	/	/	/	9,1	0,35 - 17,49	1 - 3
f85	Celery	28,27	3,5 - 100	3 - 5	7,1	0,35 - 17,49	1 - 3	7,4	0,35 - 17,49	1 - 3	12	0,69 - 49,9	2 - 4
f91	Mango	17,5	0,7 - 49,9	2 - 4	/	/	/	/	/	/	/	/	/
f92	Banana	5,83	0,35 - 17,49	1 - 3	3,3	0,35 - 17,49	1 - 3	/	/	/	12	0,69 - 49,9	2 - 4
f93	Cacao	/	/	/	/	/	/	< 0,35	0	0	/	/	/
f95	Peach	/	/	/	4,7	0,35 - 17,49	1 - 3	6,9	0,35 - 17,49	1 - 3	/	/	/
f207	Mussel	< 0,35	0	0	/	/	/	/	/	/	< 0,35	0	0
f218	Pepper	18	0,7 - 49,9	2 - 4	< 0,35	0	0	/	/	/	/	/	/
f256	Walnut	9,87	0,7 - 49,9	2 - 4	1,8	< 0,35 - 3,49	0 - 2	2,2	< 0,35 - 3,49	0 - 2	/	/	/
CCD1	Bromelain	64,17	17,5 - >100	4 - 6	/	/	/	13	0,69 - 49,9	2 - 4	23	0,69 - 49,9	2 - 4
CCD2	Horseradish peroxidase	30,2	3,5 - 100	3 - 5	/	/	/	/	/	/	/	/	/
CCD3	Ascorbat oxidase	72,43	17,5 - >100	4 - 6	/	/	/	/	/	/	/	/	/