

## Representative Reactivities Table

### QConnect ToRCH G

Positive Control

Version: 2

Date: May 16, 2018

Analyte	Manufacturer	Assay Name	Method	Unit	Mean	SD	Samples	Labs	Representative Reactivity
Rubella IgG	Abbott	Architect Rubella IgG	CMIA	IE/mL	21.89	2.15	12	2	positive
Toxo IgG	Abbott	Architect Toxo IgG	CMIA	IE/mL	8.97	0.23	12	2	positive
CMV IgG	Abbott	Architect CMV IgG	CMIA	AE/mL	40.34	3.70	12	2	positive
EBV VCA IgG	Abbott	Architect EBV VCA IgG	CMIA	S/CO	8.83	0.42	10	2	positive
EBNA IgG	Abbott	Architect EBNA-1 IgG	CMIA	S/CO	5.63	0.12	10	2	positive
Rubella IgG	DiaSorin	Liaison Rubella IgG	CLIA	IE/ml	15.78	0.57	12	2	positive
Toxo IgG	DiaSorin	Liaison Toxo IgG II	CLIA	IE/mL	18.24	2.28	12	2	positive
CMV IgG	DiaSorin	Liaison CMV IgG II	CLIA	U/mL	51.04	3.72	12	2	positive
EBV VCA IgG	DiaSorin	Liaison VCA IgG	CLIA	E/mL	20.86	2.43	12	2	cutoff
HSV-1/2 IgG	DiaSorin	Liaison HSV-1/2 IgG	CLIA	Index	11.70	1.41	12	2	positive
EBNA IgG	DiaSorin	Liaison EBNA IgG	CLIA	E/ml	31.01	3.73	12	2	positive
Rubella IgG	Siemens	Enzygnost Anti- Rubella/IgG	ELISA	IU/mL	28.00	2.26	10	2	positive
Toxo IgG	Siemens	Enzygnost Toxo/IgG	ELISA	IU/mL	18.20	3.74	10	2	positive
CMV IgG	Siemens	Enzygnost Anti-CMV/IgG	ELISA	Titer	2707	381	15	2	positive
EBV VCA IgG	Siemens	Enzygnost Anti-EBV IgG	ELISA	U/mL	< 25	-	10	2	negative
EBNA IgG	Siemens	Enzygnost Anti-EBNA/IgG	ELISA	NG U	24.73	0.69	10	2	positive
HSV-1/2 IgG	Siemens	Enzygnost HSV-1/2/IgG	ELISA	Titer	2787	292	15	2	positive

The stated mean values were obtained by testing one product batch. They show possible representative values but no target values. Test results may vary among used instruments, assays, reagent lots, and different laboratory protocols.